



Method(e):

0231

1.00–10.00 g/L O₂


20–80 %:

1.00 2.80


8.20 10.00 g/L O₂




REF 985 023




Probe / Sample
Échant. / Muestra




Schütteln / Shake
Agiter / Agitar




148 °C / 2 h



Schütteln / Shake
Agiter / Agitar



Abkühlen / Cool
Refridir / Enfriar



Säubern / Clean
Nettoyer / Limpiar



Messen / Measure
Mesurer / Medir