

# Kalium 50

## Potassium / Potasio

**Test 0-45**

REF 985 045



Rundküvettentest / Tube test


Test en cuvas rondes / Prueba en cubeta cilíndrica

2–50 mg/L K<sup>+</sup>


690 nm

Method(e) / Método


**0451**




**1 x R2**




Schütteln  
Shake  
Agiter  
Agitar




**2.0 mL** Probe  
Sample  
Echantillon  
Muestra




**10 s** Schütteln  
Shake  
Agiter  
Agitar



Säubern  
Clean  
Nettoyer  
Limpiar



**2'00 min**



Messung  
Measurement  
Mesure  
Medición