




Method(e):


0681	0682
0.003–0.460 mg/L NO <sub>2</sub> -N	0.02–1.50 mg/L NO <sub>2</sub> <sup>-</sup>

20–80 %:


0.003	0.094	0.369	0.460 mg/L NO <sub>2</sub> -N
0.02	0.32	1.20	1.50 mg/L NO <sub>2</sub> <sup>-</sup>




REF 985 068




Probe / Sample  
Échant. / Muestra




200 µL R2




Schütteln / Shake  
Agiter / Agitar



Säubern / Clean  
Nettoyer / Limpiar



10'00 min



Messen / Measure  
Mesurer / Medir