




Method(e):

0361


3–150 mg/L O₂

20–80 %:


3 32 121 150 mg/L O₂




REF 985 036




Probe / Sample
Échant. / Muestra




Schütteln / Shake
Agiter / Agitar




150 °C / 2 h



Schwenken / Invert
Brandir / Ondear



Abkühlen / Cool
Refridir / Enfriar



Säubern / Clean
Nettoyer / Limpiar



Messen / Measure
Mesurer / Medir