



Method(e):


0221
5–60 mg/L O₂

20–80 %:


5 16 49 60 mg/L O₂




REF 985 022




Probe / Sample
Échant. / Muestra




Schütteln / Shake
Agiter / Agitar




160 °C / 30 min
148 °C / 2 h



Schütteln / Shake
Agiter / Agitar



Abkühlen / Cool
Refridir / Enfriar



Säubern / Clean
Nettoyer / Limpiar



Messen / Measure
Mesurer / Medir